



TEST CERTIFICATE

ISSUED BY SIRA TEST & CERTIFICATION LIMITED

TESTS FOR THE INGRESS PROTECTION OF A DOSERAЕ PRO RADIATION MONITOR

Manufacturer: RAE Systems
Kirstinehoj 23A
2770 Kastrup
Denmark

Item tested: DoseRAE Pro radiation monitor

Standard: BS EN 60529:1992 Inc Amendments Nos 1 and 2

Deviations from Standard: None

ST&C Test Procedure: LOP 220

ST&C Test Reports: N58D30560A and 13/0271

Sample Delivery Date: 2nd April 2013

Tests Conducted On: 15th April 2013

This certificate refers to the performance of the test sample when tested against the agreed programme. It does not imply that any other samples or products necessarily comply with the requirements of the test programme.

Sira Test & Certification Limited being a UKAS accredited Test House in accordance with ISO/IEC 17025 has tested the DoseRAE Pro radiation monitor and has found it met the requirements of the Ingress Protection Code: IP 54 Category I.

S P CORK
Technical Manager

Dated 26th April 2013

This certificate may only be reproduced in its entirety and without change

Page 1 of 1

Certificate No: Sira 58D30560A
FORM 6240, Issue 8

a CSA Group Company