



DET NORSKE VERITAS
CERTIFICATION AS

*Compliance Testing & Field
Measurement*

Veritasveien 1
N-1322 HØVIK Norway
Tel: +47 67 57 99 00
Fax: +47 67 57 99 11
<http://www.dnv.com>

RAE Systems Europe
Kirstinehøj 23A

2770 Kastrup
Danmark

Att: Ole Olsen

Your ref.:
Ole Olsen

Our ref.:
dinh

Date:
2010-10-07

Statement of testing

IP67 testing of ToxiRAE 3 Toxic Gas Detectors according to IEC 60529 ED 2.1 (2001)

Hereby confirm that the following tests were performed at the Environmental testing laboratory of Det Norske Veritas Certification AS on the time period from 2010-08-04 to 2010-10-04:

- 1) Chapter 13.6 – Special conditions for first characteristic numeral 6,
- 2) Chapter 14.2.7 – Test for second characteristic numeral 7 – Temporary immersion between 0.15m and 1m.

The detectors passed the tests, see DNV test report 2010-3505 rev 02.

Yours faithfully
for DET NORSKE VERITAS CERTIFICATION AS

Einar Richter Jordfald
Head of section
Compliance Testing & Field Measurement

Huu Dung Dinh
Saksbehandler