

## IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## CB TEST CERTIFICATE

## Product

Optical Gas Imaging for Emissions Monitoring

## Name and address of the applicant

Rebellion Photonics, Inc.  
9401 Bay Area Blvd.  
#400  
Pasadena, TX, 77507  
United States of America

## Name and address of the manufacturer

Rebellion Photonics, Inc.  
9401 Bay Area Blvd.  
#400  
Pasadena, TX, 77507  
United States of America

## Name and address of the factory

*Note: When more than one factory, please report on page 2*

Rebellion Photonics, Inc.  
9401 Bay Area Blvd.  
#400  
Pasadena, TX, 77507  
United States of America

[Additional information on page 2](#)

## Ratings and principal characteristics

Facility Interface Panel (FIP): 110VAC/60Hz - 230VAC/50Hz, Max  
Current: 5A @ 110VAC, IP66  
Mini CGI Imaging Camera: 24Vdc, 0.5A, 12W Steady-State, 24W Max, IP54  
Junction Box (JB): 24Vdc, IP66  
PTZ (Security) Camera: PoE, IP67, 17.1W Max

## Trademark (if any)



Honeywell

## Customer's Testing Facility (CTF) Stage used

## Model / Type Ref.

Mini Gas Cloud Imaging System - MNC Series

## Additional information (if necessary may also be reported on page 2)

[Additional information on page 2](#)

## A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010

**Comments:**

Refer to the test report for the summary of compliance with National Deviations.

## As shown in the Test Report Ref. No. which forms part of this Certificate

104450631DAL-001

This CB Test Certificate is issued by the National Certification Body

Intertek Testing Services NA, Inc.  
545 E. Algonquin Road  
Suite F  
Arlington Heights, IL 60005  
United States of America



Date: 2022-04-18

Signature: John Quigley